

## EUVSEC

**Bakterier:** *Enterobacteriaceae, Pseudomonas*  
**Medium:** MH  
**Volumen:** 50 µl pr brønd

	1	2	3	4	5	6	7	8	9	10	11	12
A	SMX 1024	TMP 32	CIP 8	TET 64	MERO 16	AZI 64	NAL 128	CHL 128	TGC 8	COL 16	AMP 64	GEN 32
B	SMX 512	TMP 16	CIP 4	TET 32	MERO 8	AZI 32	NAL 64	CHL 64	TGC 4	COL 8	AMP 32	GEN 16
C	SMX 256	TMP 8	CIP 2	TET 16	MERO 4	AZI 16	NAL 32	CHL 32	TGC 2	COL 4	AMP 16	GEN 8
D	SMX 128	TMP 4	CIP 1	TET 8	MERO 2	AZI 8	NAL 16	CHL 16	TGC 1	COL 2	AMP 8	GEN 4
E	SMX 64	TMP 2	CIP 0.5	TET 4	MERO 1	AZI 4	NAL 8	CHL 8	TGC 0.5	COL 1	AMP 4	GEN 2
F	SMX 32	TMP 1	CIP 0.25	TET 2	MERO 0.5	AZI 2	NAL 4	FOT 1	TGC 0.25	TAZ 2	AMP 2	GEN 1
G	SMX 16	TMP 0.5	CIP 0.12	CIP 0.03	MERO 0.25	MERO 0.06	FOT 4	FOT 0.5	TAZ 8	TAZ 1	AMP 1	GEN 0.5
H	SMX 8	TMP 0.25	CIP 0.06	CIP 0.015	MERO 0.12	MERO 0.03	FOT 2	FOT 0.25	TAZ 4	TAZ 0.5	POS KON	POS KON

Kode	Antimikrobielt stof (14)	Testinterval (µg/ml)
AMP	AMPICILLIN	1-64
AZI	AZITHROMYCIN	2-64
FOT	CEFOTAXIME	0.25-4
CHL	CHLORAMPHENICOL	8-128
CIP	CIPROFLOXACIN	0.015-8
COL	COLISTIN	1-16
GEN	GENTAMICIN	0.5-32
MERO	MEROPENEM	0.03-16
NAL	NALIDIXAN	4-128
SMX	SULPHAMETHOXAZOLE	8-1024
TAZ	CEFTAZIDIME	0.5-8
TET	TETRACYKLIN	2-64
TGC	TIGECYCLINE	0.25-8
TMP	TRIMETHOPRIM	0.25-32

## EUVSEC2

**Bakterier:** ESBL suspekter isolater (resistens overfor FOT, TAZ eller MERO i EUVSEC, primært *E. coli* og salmonella)

**Medium:** MH

**Volumen:** 50 µl pr. brønd

	1	2	3	4	5	6	7	8	9	10	11	12
<b>A</b>	FOX 0.5	FOX 1	FOX 2	FOX 4	FOX 8	FOX 16	FOX 32	FOX 64	FOT 0.25	FOT 0.5	FOT 1	TRM 128
<b>B</b>	ETP 0.015	ETP 0.03	ETP 0.06	ETP 0.12	ETP 0.25	ETP 0.5	ETP 1	ETP 2	FOT 2	FOT 4	FOT 8	TRM 64
<b>C</b>	IMI 0.12	IMI 0.25	IMI 0.5	IMI 1	IMI 2	IMI 4	IMI 8	IMI 16	FOT 16	FOT 32	FOT 64	TRM 32
<b>D</b>	MERO 0.03	MERO 0.06	MERO 0.12	MERO 0.25	MERO 0.5	MERO 1	MERO 2	MERO 4	MERO 8	MERO 16	TRM 2	TRM 16
<b>E</b>	TAZ 0.25	TAZ 0.5	TAZ 1	TAZ 2	TAZ 4	TAZ 8	TAZ 16	TAZ 32	TAZ 64	TAZ 128	TRM 1	TRM 8
<b>F</b>	FEP 0.06	FEP 0.12	FEP 0.25	FEP 0.5	FEP 1	FEP 2	FEP 4	FEP 8	FEP 16	FEP 32	TRM 0.5	TRM 4
<b>G</b>	F/C 0.06/4	F/C 0.12/4	F/C 0.25/4	F/C 0.5/	F/C 1/4	F/C 2/4	F/C 4/4	F/C 8/4	F/C 16/4	F/C 32/4	F/C 64/4	POS KON
<b>H</b>	T/C 0.12/4	T/C 0.25/4	T/C 0.5/4	T/C 1/4	T/C 2/4	T/C 4/4	T/C 8/4	T/C 16/4	T/C 32/4	T/C 64/4	T/C 128/4	POS KON

<b>Kode</b>	<b>Antimikrobielt stof (10)</b>	<b>Testinterval (µg/ml)</b>
ETP	ERTAPENEM	0.015-2
FEP	CEFEPIME (4-gen)	0.06 - 32
FOT	CEFOTAXIME (3-gen)	0.25 – 64
F/C	CEFOTAXIME / CLAVULANIC ACID	0.06/4 - 64/4
FOX	CEFOXITIN (2-gen og/eller cephamycin)	0.5 - 64
IMI	IMIPENEM (carbapenem)	0.12 - 16
MERO	MEROPENEM (carbapenem)	0.03 - 16
TAZ	CEFTAZIDIME (3-gen)	0.25 - 128
T/C	CEFTAZIDIME / CLAVULANIC ACID	0.12/4 - 128/4
TRM	TERMOCILIN	0.5-64

## EUVENC

**Bakterier:** Enterokokker  
**Medium:** MH  
**Volumen:** 50 µl pr. brønd

	1	2	3	4	5	6	7	8	9	10	11	12
A	VAN 128	TEI 64	SYN 64	TET 128	DAP 32	CIP 16	ERY 128	TGC 4	LZD 64	GEN 1024	AMP 64	CHL 128
B	VAN 64	TEI 32	SYN 32	TET 64	DAP 16	CIP 8	ERY 64	TGC 2	LZD 32	GEN 512	AMP 32	CHL 64
C	VAN 32	TEI 16	SYN 16	TET 32	DAP 8	CIP 4	ERY 32	TGC 1	LZD 16	GEN 256	AMP 16	CHL 32
D	VAN 16	TEI 8	SYN 8	TET 16	DAP 4	CIP 2	ERY 16	TGC 0.5	LZD 8	GEN 128	AMP 8	CHL 16
E	VAN 8	TEI 4	SYN 4	TET 8	DAP 2	CIP 1	ERY 8	TGC 0.25	LZD 4	GEN 64	AMP 4	CHL 8
F	VAN 4	TEI 2	SYN 2	TET 4	DAP 1	CIP 0.5	ERY 4	TGC 0.12	LZD 2	GEN 32	AMP 2	CHL 4
G	VAN 2	TEI 1	SYN 1	TET 2	DAP 0.5	CIP 0.25	ERY 2	TGC 0.06	LZD 1	GEN 16	AMP 1	POS KON
H	VAN 1	TEI 0.5	SYN 0.5	TET 1	DAP 0.25	CIP 0.12	ERY 1	TGC 0.03	LZD 0.5	GEN 8	AMP 0.5	POS KON

Kode	Antimikrobielt stof (12)	Testinterval (µg/ml)
AMP	AMPICILLIN	0.5-64
CHL	CHLORAMPHENICOL	4-128
CIP	CIPROFLOXACIN	0,12-16
DAP	DAPTOMYCIN	0.25-32
ERY	ERYTHROMYCIN	1-128
GEN	GENTAMICIN	8-1024
LZD	LINEZOLID	0,5-64
SYN	QUINUPRISTIN/DALFOPRISTIN	0,5-64
TEI	TEICOPLANIN	0,5-64
TET	TETRACYCLINE	1-128
TGC	TIGECYCLINE	0,03-4
VAN	VANCOMYCIN	1-128

## EUCAMP2

**Bakterier:** *Campylobacter*  
**Medium:** MH+blod  
**Volumen:** 100µl pr brønd

	1	2	3	4	5	6	7	8	9	10	11	12
A	ERY 128	CIP 16	TET 64	GEN 16	NAL 64	STR 16	ERY 128	CIP 16	TET 64	GEN 16	NAL 64	STR 16
B	ERY 64	CIP 8	TET 32	GEN 8	NAL 32	STR 8	ERY 64	CIP 8	TET 32	GEN 8	NAL 32	STR 8
C	ERY 32	CIP 4	TET 16	GEN 4	NAL 16	STR 4	ERY 32	CIP 4	TET 16	GEN 4	NAL 16	STR 4
D	ERY 16	CIP 2	TET 8	GEN 2	NAL 8	STR 2	ERY 16	CIP 2	TET 8	GEN 2	NAL 8	STR 2
E	ERY 8	CIP 1	TET 4	GEN 1	NAL 4	STR 1	ERY 8	CIP 1	TET 4	GEN 1	NAL 4	STR 1
F	ERY 4	CIP 0.5	TET 2	GEN 0.5	NAL 2	STR 0.5	ERY 4	CIP 0.5	TET 2	GEN 0.5	NAL 2	STR 0.5
G	ERY 2	CIP 0.25	TET 1	GEN 0.25	NAL 1	STR 0.25	ERY 2	CIP 0.25	TET 1	GEN 0.25	NAL 1	STR 0.25
H	ERY 1	CIP 0.125	TET 0.5	GEN 0.125	POS KON	POS KON	ERY 1	CIP 0.125	TET 0.5	GEN 0.125	POS KON	POS KON

Kode	Antimikrobielt stof (7)	Testinterval (µg/ml)
CIP	CIPROFLOXACIN	0.125-16
ERY	ERYTHROMYCIN	1-128
GEN	GENTAMICIN	0.125-16
NAL	NALIDIXIC ACID	1-64
STR	STREPTOMYCIN	0.25-16
TET	TETRACYKLIN	0.5-64

## DKMVN4

**Bakterier:** *Enterobacteriaceae, Pseudomonas*

**Medium:** MH

**Volumen:** 50 µl pr brønd

	1	2	3	4	5	6	7	8	9	10	11	12
A	CHL 64	FFN 64	TMP 32	AUG 32	XNL 8	NAL 64	NEO 32	SPE 256	STR 128	TET 32	SMX 1024	COL 16
B	CHL 32	FFN 32	TMP 16	AUG 16	XNL 4	NAL 32	NEO 16	SPE 128	STR 64	TET 16	SMX 512	COL 8
C	CHL 16	FFN 16	TMP 8	AUG 8	XNL 2	NAL 16	NEO 8	SPE 64	STR 32	TET 8	SMX 256	COL 4
D	CHL 8	FFN 8	TMP 4	AUG 4	XNL 1	NAL 8	NEO 4	SPE 32	STR 16	TET 4	SMX 128	COL 2
E	CHL 4	FFN 4	TMP 2	AUG 2	XNL 0.5	NAL 4	NEO 2	SPE 16	STR 8	TET 2	SMX 64	COL 1
F	CHL 2	FFN 2	TMP 1	CIP 0.015	CIP 0.03	CIP 0.06	CIP 0.125	CIP 0.25	CIP 0.5	CIP 1	CIP 2	CIP 4
G	AMP 1	AMP 2	AMP 4	AMP 8	AMP 16	AMP 32	GEN 0.5	GEN 1	GEN 2	GEN 4	GEN 8	GEN 16
H	FOT 0.125	FOT 0.25	FOT 0.5	FOT 1	FOT 2	FOT 4	APR 4	APR 8	APR 16	APR 32	POS KON	POS KON

Kode	Antimikrobielt stof (17)	Testinterval (µg/ml)
AUG2	AMOXICILLIN+CLAVULANAT (2:1)	2/1-32/16
AMP	AMPICILLIN	1-32
APR	APRAMYCIN	4-32
FOT	CEFOTAXIME	0.125-4
XNL	CEFTIOFUR	0.5-8
CHL	CHLORAMPHENICOL	2-64
CIP	CIPROFLOXACIN	0.015-4
COL	COLISTIN	1-16
FFN	FLORFENICOL	2-64
GEN	GENTAMICIN	0.5-16
NAL	NALIDIXAN	4-64
NEO	NEOMYCIN	2-32
SPE	SPECTINOMYCIN	16-256
STR	STREPTOMYCIN	8-128
SMX	SULPHAMETHOXAZOLE	64-1024
TET	TETRACYKLIN	2-32
TMP	TRIMETHOPRIM	1-32

## DKVP

**Bakterier:** Stafylokokker / streptokokker / *Listeria*  
**Medium:** MH (stafylokok), MH+blod (streptokok, *Listeria*)  
**Volumen:** 50µl pr. brønd

	1	2	3	4	5	6	7	8	9	10	11	12
A	PEN 0.06	PEN 0.12	PEN 0.25	PEN 0.5	PEN 1	PEN 2	PEN 4	PEN 8	PEN 16	TMP 32	SXT 16	SMX 512
B	TIA 0.25	TIA 0.5	TIA 1	TIA 2	TIA 4	TIA 8	TIA 16	TIA 32	CHL 64	TMP 16	SXT 8	SMX 256
C	ERY 0.25	ERY 0.5	ERY 1	ERY 2	ERY 4	ERY 8	ERY 16	SPE 256	CHL 32	TMP 8	SXT 4	SMX 128
D	TET 0.5	TET 1	TET 2	TET 4	TET 8	TET 16	TET 32	SPE 128	CHL 16	TMP 4	SXT 2	SMX 64
E	CIP 0.12	CIP 0.25	CIP 0.5	CIP 1	CIP 2	CIP 4	CIP 8	SPE 64	CHL 8	TMP 2	SXT 1	SMX 32
F	GEN 0.25	GEN 0.5	GEN 1	GEN 2	GEN 4	GEN 8	GEN 16	SPE 32	CHL 4	TMP 1	SXT 0.5	POS KON
G	FOX 0.5	FOX 1	FOX 2	FOX 4	FOX 8	FOX 16	FOX 32	SPE 16	CHL 2	TMP 0.5	SXT 0.25	POS KON
H	FFN 1	FFN 2	FFN 4	FFN 8	FFN 16	FFN 32	FFN 64	STR 4	STR 8	STR 16	STR 32	STR 64

Kode	Antimikrobielt stof (14)	Testinterval (µg/ml)
FOX	CEFOXITIN	0.5-32
CHL	CHLORAMPHENICOL	2-64
CIP	CIPROFLOXACIN	0.12-8
ERY	ERYTHROMYCIN	0.25-16
FFN	FLORFENICOL	1-64
GEN	GENTAMICIN	0.25-16
PEN	PENICILLIN	0.06-16
SPE	SPECTINOMYCIN	16-256
STR	STREPTOMYCIN	4-64
SMX	SULPHAMETHOXAZOL	32-512
TET	TETRACYKLIN	0.5-32
TIA	TIAMULIN	0.25-32
TMP	TRIMETHOPRIM	0.5-32
SXT	TMP+SMX	0.25-16

## DKVPE

**Bakterie:** Enterokokker (primært *E. faecalis* og *E. faecium*)  
**Bouillon:** MH  
**Volumen:** 50 µl pr. brønd

	1	2	3	4	5	6	7	8	9	10	11	12
A	TGC 0.015	TGC 0.03	TGC 0.06	TGC 0.12	TGC 0.25	TGC 0.5	TGC 1	AMP 2	AMP 4	AMP 8	AMP 16	AMP 32
B	ERY 0.5	ERY 1	ERY 2	ERY 4	ERY 8	ERY 16	ERY 32	PEN 2	PEN 4	PEN 8	PEN 16	PEN 32
C	GEN 16	GEN 32	GEN 64	GEN 128	GEN 256	GEN 512	GEN 1024	KAN 128	KAN 256	KAN 512	KAN 1024	KAN 2048
D	SYN 0.25	SYN 0.5	SYN 1	SYN 2	SYN 4	SYN 8	SYN 16	LZD 0.5	LZD 1	LZD 2	LZD 4	LZD 8
E	CHL 2	CHL 4	CHL 8	CHL 16	CHL 32	CHL 64	STR 64	STR 128	STR 256	STR 512	STR 1024	STR 2048
F	VAN 1	VAN 2	VAN 4	VAN 8	VAN 16	VAN 32	TET 1	TET 2	TET 4	TET 8	TET 16	TET 32
G	SAL 2	SAL 4	SAL 8	SAL 16	TEI 0.25	TEI 0.5	TEI 1	TEI 2	TEI 4	TEI 8	TEI 16	TEI 32
H	CIP 0.5	CIP 1	CIP 2	CIP 4	CIP 8	CIP 16	CIP 32	CIP 64	CIP 128	CIP 256	POS KON	POS KON

### Kode      Antimikrobielt stof (15)      Testinterval (µg/ml)

AMP	AMPICILLIN	2-32
CHL	CHLORAMPHENICOL	2-64
CIP	CIPROFLOXACIN	0,5-256
ERY	ERYTHROMYCIN	0,5-32
GEN	GENTAMICIN	16-1024
KAN	KANAMYCIN	128-2048
LZD	LINEZOLID	0,5-8
PEN	PENICILLIN	2-32
SAL	SALINOMYCIN	2-16
STR	STREPTOMYCIN	64-2048
SYN	SYNERCID	0,25-16
TEI	TEICOPLANIN	0,25-32
TET	TETRACYCLINE	1-32
TGC	TIGECYCLINE	0,015-1
VAN	VANCOMYCIN	1-32

## SENSITITRE CUSTOM PLATE FORMAT

0801161148

Plate Code: **FRCOL**

Date: **13-Jan-16**

	1	2	3	4	5	6	7	8	9	10	11	12
<b>A</b>	COL	COL	COL	COL	COL	COL	COL	COL	COL	COL	COL	POS
	0.12	0.25	0.5	1	2	4	8	16	32	64	128	CON
<b>B</b>	COL	COL	COL	COL	COL	COL	COL	COL	COL	COL	COL	POS
	0.12	0.25	0.5	1	2	4	8	16	32	64	128	CON
<b>C</b>	COL	COL	COL	COL	COL	COL	COL	COL	COL	COL	COL	POS
	0.12	0.25	0.5	1	2	4	8	16	32	64	128	CON
<b>D</b>	COL	COL	COL	COL	COL	COL	COL	COL	COL	COL	COL	POS
	0.12	0.25	0.5	1	2	4	8	16	32	64	128	CON
<b>E</b>	COL	COL	COL	COL	COL	COL	COL	COL	COL	COL	COL	POS
	0.12	0.25	0.5	1	2	4	8	16	32	64	128	CON
<b>F</b>	COL	COL	COL	COL	COL	COL	COL	COL	COL	COL	COL	POS
	0.12	0.25	0.5	1	2	4	8	16	32	64	128	CON
<b>G</b>	COL	COL	COL	COL	COL	COL	COL	COL	COL	COL	COL	POS
	0.12	0.25	0.5	1	2	4	8	16	32	64	128	CON
<b>H</b>	COL	COL	COL	COL	COL	COL	COL	COL	COL	COL	COL	POS
	0.12	0.25	0.5	1	2	4	8	16	32	64	128	CON

**ANTIMICROBICS**

COL Colistin  
 POS Positive Control

READ METHOD: MANUAL

ORGANISMS TO BE TESTED IN THE PLATE: GRAM NEGATIVE

SUBSTRATE IN WELLS: NOT REQUIRED

RECONSTITUTION VOLUME: 50µl

INOCULUM CONCENTRATION: 1 x 10<sup>5</sup> cfu/ml

PRODUCT LABELLING: IN-VITRO DIAGNOSTIC USE ONLY

LABEL COLOUR: RED

PLATE EXPIRY: 24 Months

PACK INSERT: To be assigned