



DTU Food
National Food Institute

EQAS 2014

Salmonella and E. coli

EURL-AR workshop, April 23-24th, 2015

Participation

Salmonella

32 countries

35 sets of *Salmonella* results

One set excluded due to lack of full-range dilution tests

E. coli


31 countries

35 sets of *E. coli* results included


2 sets excluded due to DD data



Comparison to former EQASs – *E. coli*



E. coli results – pr. lab



One participant (lab #58) had 6.8 % deviations



ESBL-producing test strains

ESBL-detection is mandatory in the *E. coli* EQAS


Percentage reporting correct ESBL-results		EC 8.2	EC 8.3	EC 8.4	EC 8.7
ESBL-genes harboured in the test strain		KPC-2	CTX-M-1	CMY-2	OXA-48
ESBL-, pAmpC- and carbapenemase-producing strain – expected results		carbapenemase	ESBL	AmpC	carbapenemase
Obtained results	Presumptive ESBL-producer		34/36 (94.4%)	-	
	Presumptive pAmpC-producer	2/36 (5.6%)	1/36 (2.8%)	32/36 (88.9%)	
	Presumptive ESBL+pAmpC	4/36 (11.1%)	1/36 (2.8%)	4/36 (11.1%)	-
	Presumptive carbapenemase-producer	21/36 (58.3%)	-	-	16/36 (44.4%)
	unusual phenotype	9/36 (25%)	-	-	3/36 (8.3%)
	Not ESBL-, pAmpC- or carbapenemase-producing	-	-	-	17/36 (47.2%)



Labs #54 did not provide results for all strains


EC 8.7/meropenem

54.8% deviations in panel 1 and 35% in panel 2




Participants had slightly better results in second panel; probably due to considering the results in relation to other carbapenems in the panel. Expected MIC value: 0, 25 mg/l ; R

EC 8.7 meropenem both panels total deviation 47%




Because the issues with this strain/antimicrobial combination are related to the OXA-48 gene phenotype the data was kept in the report. Expected MIC value: 0, 25 mg/l ; R

EC 8.7/imipenem total deviation 42.1%




The issues with this strain/ antimicrobial combination are related to the expected OXA-48 gene phenotype, therefore the data was kept in the report. Expected MIC value: 1 mg/l ; R

Comparison to former EQASs - *Salmonella*



Salmonella results – pr. lab



ESBL-producing test strains


ESBL-detection is mandatory in the *Salmonella* EQAS

	Strain S-9.3	Strain S-9.4	Strain S-9.5	Strain S-9.6	
ESC-genes harboured in the test strain	<i>bla</i> _{CMY-2}	<i>bla</i> _{CTX-M-9} <i>bla</i> _{TEM-1}	<i>bla</i> _{VIM-2} <i>bla</i> _{TEM-1}	<i>bla</i> _{OXA-48}	
ESBL-, pAmpC- and carbapenemase-producing strain – expected results	pAmpC	ESBL	carbapenemase	carbapenemase	
Obtained results	Confirmed ESBL-producer	-	32/35 (91%)	-	
	Confirmed ESBL + pAmpC-producer	3/35 (9%)	-	-	
	Confirmed pAmpC-producer	32/35 (91%)	-	4/35 (11%)	-
	Confirmed carbapenemase-producer	-	-	28/35 (80%)	29/35 (83%)
	Confirmed unusual phenotype	-	2/35 (6%)	3/35 (9%)	4/35 (11%)
	Not ESBL-, pAmpC- or carbapenemase-producing	-	1/35 (3%)	-	2/35 (6%)

Marked in grey: incorrect results




S9.8/colistin excluded 36% deviations




High level of deviation could not be explained
=> results excluded from evaluation
Expected MIC value: 2 mg/l ; S

S-9.5/meropenem (not excluded)
36% deviations in panel 1 and 30% in panel 2



High level of deviation
Meropenem => results not excluded from evaluation
Expected MIC value: 1 mg/l ; R

S-9.7/tigecycline (not excluded) 29% deviations



High level of deviation.

New antimicrobial => results not excluded from evaluation

Expected MIC value: 2 mg/l ; R

NRL-AR - Czech Republic - point of view

Our results of problematic strains and antibiotic combination :

- **EC 8.7** MER 0,25 mg/l (P1,P2); IMP 1 mg/l – fits to expected
- **S 9.5** MER 0,25mg/l (P1,P2) – expected (1; Rmg/l)
- **S 9.7** TGC 1 mg/l (P1) – expected (2; R mg/l)
Result fits with ISO 20776 – one dilution step below only
- **S 9.8** COL 4 mg/l (P1) – expected (2; S mg/l) –
Result fits with ISO 20776 – one dilution step up only

NRL- AR - Czech Republic - point of view

Analysis of the possible causes of results deviation:

Suspensions solutions (demineralised water, CAMHB)

- Home made
- Original brand (TREK)

Primary inoculum density - 0,5 McF (1×10^8 CFU/ml)

- Calibrated densitometer - TRECK 0,5 McF density control
- Spectrometric control - 0,08-0,13 (625 nm)

Volume of suspension which was transferred into the CAMHB

- Manufacturer instruction 10 μ l (1×10^5 CFU/ml)
- EURL protocol 50 μ l (5×10^5 CFU/ml)
- ISO 20776-1 [5×10^5 CFU/ml ($2-8 \times 10^5$)]

Major hazard point

Final transfer of 50 μ l CAMHB into the MIC plate wells

- Multichannel pipette inoculation
- SENSITITRE autoinoculator

Quality control with the E. coli ATCC 25922 strain

Major hazard point

- Lowest and highest ranges of antimicrobial MIC which is valid for QCS
- Lowest and highest dilution range of antimicrobials to the MIC plate

NRL-AR - Czech Republic - point of view

Analysis of the possible causes of results deviation:

Final density of inoculum in the wells: 5×10^5 CFU/ml ($2-8 \times 10^5$):

- Semiquantitative checking 10µl from PGC >> 9.9 ml saline >> 100 µl cultivate onto the plate >> grow 20-80 colonies (ISO 20776-1,CLSI)
- Quantitative checking by ISO 4833-1 (direct number of CFU)

Our opinion:

- Transferring 10 µl of 0.5 McF suspension into CAMHB only, may be resulting of the lower final concentrations ($9 \times 10^4 - 2 \times 10^5$)

Quality control check by the E. coli ATCC 25922:

- Lowest dilution of antimicrobials in the EUVSEC plate does not fit with the lowest QCS range of some antimicrobials

Our opinion:

- In some cases we can not check of lowest QCS border of antimicrobials

		MIC (mg/l)											
Range		0,008	0,015	0,03	0,06	0,12	0,25	0,5	1	2	4	8	16
MER	QCS												
	PLATE												
TGC	QCS												
	PLATE												
COL	QCS												
	PLATE												

NRL-AR - Czech Republic - conclusion

System error input material or workflow >>>system error of achieved MIC result:

- Lowest inoculum density >>> lowest MIC (MER, TGC may correspond)
- But our result of COL (higher than expected) defies of this rule

Our deviations in described cases are not alone:

- MER 36%, TGC 29%, COL 36%

There may be other unmentioned reasons for such a high percentage deviations appointed by antimicrobials ???

- Variability of antimicrobial activity of the bacterial population of the test samples in time and space
- Another unnamed and unexpected influences (reasons)

DISCUSSION